

distributed in the US/Canada by:  
**EAGLE BIOSCIENCES, INC.**  
 20A NW Blvd, Suite 112 Nashua, NH 03063  
 Phone: 617-419-2019 FAX: 617-419-1110  
[www.EagleBio.com](http://www.EagleBio.com) info@eaglebio.com



# Anti-Daratumumab (Darzalex™) ELISA

**REF** : KBI2028

Ver 1.1

**RUO**

Enzyme Immunoassay for the quantitative determination of Anti-Daratumumab in serum, plasma and cell culture supernatant

<b>RUO</b>	For Research Use	<b>REF</b>	Catalog Number
	Store At	<b>LOT</b>	Batch Code
	Manufactured By		Biological Risk
	Expiry Date		Consult Operating Instructions

*For Research Use Only. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of KRISHGEN BioSystems is strictly prohibited.*

**REF** KBI2028

96 tests



**KRISHGEN BioSystems**

| For US / Europe: toll free +1(888)-970-0827 tel: +1(562)-568-5005  
 For Asia / India: tel: +91(22)-49198700  
 Email: sales@krishgen.com

**Introduction:**

Daratumumab (trade name Darzalex) is an anti-cancer drug. It binds to CD38, which multiple myeloma cells overexpress. Daratumumab was originally developed by Genmab, but it is now being jointly developed by Genmab along with the Johnson & Johnson subsidiary Janssen Biotech, which acquired worldwide commercialization rights to the drug from Genmab.

Daratumumab was given breakthrough therapy drug status in 2013 for multiple myeloma. It was awarded orphan drug status for multiple myeloma, diffuse large B cell lymphoma, follicular lymphoma, and mantle cell lymphoma. Anti-Drug Antibodies (ADA) may induce unwanted side effects in biopharmaceuticals.


Hence, ADA has been subjected to increase in scrutiny by the regulatory authorities using immunogenicity safety studies. ADA has been observed in pre-clinical and clinical studies, resulting in significant changes in toxicology, pharmacokinetics and efficacy. These effects result from the generation of drug-induced (neutralizing) autoantibodies against Daratumumab and can be responsible for allergic reaction, or even anaphylactic shock. This ELISA kit detects antibodies for Anti-Daratumumab and may be used for monitoring immunogenicity.

**Intended Use:**

The Anti-Daratumumab (Darzalex™) ELISA is used as an analytical tool for quantitative determination of Anti-Daratumumab in serum, plasma and cell culture supernatant.

**Principle:**

The method employs the quantitative sandwich enzyme immunoassay technique. Daratumumab is pre-coated onto microwells. Samples and standards are pipetted into microwells and antibodies to Daratumumab present in the sample are bound by the capture antibody. Then, a HRP (horseradish peroxidase) conjugated Daratumumab is pipetted and incubated. After washing microwells in order to remove any non-specific binding, the ready to use substrate solution (TMB) is added to microwells and color develops proportionally to the amount of Anti-Daratumumab in the sample. Color development is then stopped by addition of stop solution. Absorbance is measured at 450 nm.



**Materials Provided:**

1. Daratumumab Coated Microtiter Plate (12 x 8 wells) - 1 no
2. Anti-Daratumumab Standard, (0.5 ml/vial) - 0, 10, 20, 40, 80, 160, 320 and 640 ng/ml
3. Daratumumab:HRP Conjugate - 12 ml
4. Assay Diluent - 6 ml
5. Sample Diluent - 50 ml
6. (20X) Wash Buffer - 25 ml
7. TMB Substrate - 12 ml
8. Stop Solution - 12 ml
9. Instruction Manual

**Materials to be provided by the End-User:**

1. Microtiter Plate Reader able to measure absorbance at 450 nm.
2. Adjustable pipettes and multichannel pipettor to measure volumes ranging from 25 ul to 1000 ul
3. Deionized (DI) water

4. Wash bottle or automated microplate washer
5. Graph paper or software for data analysis
6. Timer
7. Absorbent Paper

**Handling/Storage:**

1. All reagents should be stored at 2°C to 8°C for stability.
2. All the reagents and wash solutions should be used within 12 months from manufacturing date.
3. Before using, bring all components to room temperature (18-25°C). Upon assay completion ensure all components of the kit are returned to appropriate storage conditions.
4. The Substrate is light-sensitive and should be protected from direct sunlight or UV sources.

**Health Hazard Warnings:**

1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin.
2. For Research Use Only.

**Sample Preparation and Storage:**

Blood is taken by venipuncture. Serum is separated after clotting by centrifugation. Plasma can be used, too. Lipaemic, hemolytic or contaminated samples should not be run. Repeated freezing and thawing should be avoided. If samples are to be used for several assays, initially aliquot samples and keep at -20°C.

For Cell Culture Supernatant - If necessary, centrifuge to remove debris prior to analysis. Samples can be stored at -20°C or -80°C. Avoid repeated freeze-thaw cycles.

**Preparation Before Use:**

Allow samples to reach room temperature prior to assay. Take care to agitate patient samples gently in order to ensure homogeneity.

Serum and Plasma Test Sample preparation - Samples have to be diluted 1:10 to 1:100 (v/v), e.g. for 1:100 (5 ul sample + 495 ul sample diluent) prior to assay. The samples may be kept at 2 - 8°C for up to three days. Long-term storage requires -20°C.

**Reagent Preparation (all reagents should be diluted immediately prior to use):**

1. Label any aliquots made with the kit Lot No and Expiration date and store it at appropriate conditions mentioned.
2. Bring all reagents to Room temperature before use.
3. To make Wash Buffer (1X); dilute 25 ml of 20X Wash Buffer in 475 ml of DI water.

**Procedural Notes:**

1. In order to achieve good assay reproducibility and sensitivity, proper washing of the plates to remove excess un-reacted reagents is essential.
2. High Dose Hook Effect may be observed in samples with very high concentrations of Anti-Daratumumab. High Dose Hook Effect is due to excess of antibody for very high concentrations of Anti-Daratumumab present in the sample. High Dose Hook effect is most likely encountered from samples early in the purification process. If Hook Effect is possible, the samples to be assayed should be diluted with a compatible diluent. Thus if the Anti-Daratumumab concentration of the undiluted sample is less than the diluted sample, this may be indicative of the Hook Effect.
3. Avoid assay of Samples containing sodium azide (NaN<sub>3</sub>), as it could destroy the HRP activity resulting in under-estimation of the amount of Anti-Daratumumab.

4. It is recommended that all Standards and Samples be assayed in duplicates.
5. Maintain a repetitive timing sequence from well to well for all the steps to ensure that the incubation timings are same for each well.
6. If the Substrate has a distinct blue color prior to use it may have been contaminated and use of such substrate can lead to compromisation of the sensitivity of the assay.
7. The plates should be read within 30 minutes after adding the Stop Solution.
8. Make a work list in order to identify the location of Standards and Samples.

#### Assay Procedure:

1. It is strongly recommended that all Controls and Samples be run in duplicates. A standard curve is required for each assay. All steps must be performed at 37°C.
2. Add **100 ul** of **Standards** or **Samples** into the respective wells.
3. Cover the plate and incubate for 60 minutes at 37°C.
4. Aspirate and wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe of any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step.
5. Pipette without delay in the same order **100 ul** of **Daratumumab: HRP Conjugate** into each well.
6. Cover the plate and incubate for 60 minutes at 37°C.
7. Aspirate and wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe of any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step.
8. Add **100 ul** of **TMB Substrate** in each well.
9. Incubate the plate at **37°C** for 30 minutes in dark. DO NOT SHAKE or else it may result in higher backgrounds and worse precision. Positive wells should turn bluish in color.
10. Pipette out **100 ul** of **Stop Solution**. Wells should turn from blue to yellow in color.
11. Read the absorbance at 450 nm with a microplate reader.

#### Calculation of Results:

Determine the Mean Absorbance for each set of duplicate or triplicate Standards and Samples. Using graph paper, plot the average value (absorbance 450 nm) of each standard on the Y-axis versus the corresponding concentration of the standards on the X-axis. Draw the best fit curve through the standard points. To determine the unknown Anti- Daratumumab concentrations, find the unknown's Mean Absorbance value on the Y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the X-axis and read the Anti-Daratumumab Concentration. If samples were diluted, multiply by the appropriate dilution factor.


Software which is able to generate a cubic spline curve-fit or a polynomial curve (2<sup>nd</sup> order) is best recommended for automated results.

#### Note:

It is recommended to repeat the assay at a different dilution factor in the following cases:

- If the sample absorbance value is below the first standard.
- If the absorbance value is equivalent or higher than the 640 ng/ml standard.

**Typical Graph**



**Quality Control:**

It is recommended that for each laboratory assay appropriate quality control samples in each run to be used to ensure that all reagents and procedures are correct.

**Performance Characteristics of the Kit:**

This kit has been validated as per EMA/FDA guidelines in line with ICH Code for Harmonization of Biological Assays.

**Sensitivity:**

**Limit Of Detection:** It is defined as the lowest detectable concentration corresponding to a signal of Mean of '0' standard plus 2\* SD. 10 replicates of '0' standards were evaluated and the LOD was found to be less than 10 ng/ml

**Specificity:**

The antibodies used in the kit are monoclonal antibodies, anti-idiotypic and specific for Anti-Daratumumab. The calibrators / standards used are calibrated against commercially sourced (Darzalex™).

**Linearity:**

Standards provided in the kit will be used for measuring the linearity range of Anti-Daratumumab present in matrix.

**Precision:**

Precision is defined as the percent coefficient of variation (%CV) i.e. standard deviation divided by the mean and multiplied by 100. Assay precision was determined by both intra (n=5 assays) and inter assay (n=5 assays) reproducibility on two pools with low (10 ng/ml), medium (80 ng/ml) and high (640 ng/ml) concentrations.

While actual precision may vary from laboratory to laboratory and technician to technician, it is recommended that all operators achieve precision below these design goals before reporting results.

Pool	Intra Assay %CV	Inter Assay %CV
Low	<10%	<10%
Medium	<5%	<5%
High	<5%	<5%

**Safety Precautions:**

- **This kit is for Research Use only.** Follow the working instructions carefully.
- The expiration dates stated on the kit are to be observed. The same relates to the stability stated for reagents
- Do not use or mix reagents from different lots.
- Do not use reagents from other manufacturers.
- Avoid time shift during pipetting of reagents.

- All reagents should be kept in the original shipping container.
- Some of the reagents contain small amount of sodium azide (< 0.1 % w/w) as preservative. They must not be swallowed or allowed to come into contact with skin or mucosa.
- Source materials maybe derived from human body fluids or organs used in the preparation of this kit were tested and found negative for HBsAg and HIV as well as for HCV antibodies. However, no known test guarantees the absence of such viral agents. Therefore, handle all components and all patient samples as if potentially hazardous.
- Since the kit contains potentially hazardous materials, the following precautions should be observed
  - Do not smoke, eat or drink while handling kit material
  - Always use protective gloves
  - Never pipette material by mouth
  - Wipe up spills promptly, washing the affected surface thoroughly with a decontaminant.
- In any case GLP should be applied with all general and individual regulations to the use of this kit.



**References:**

Pharmacokinetics of Daratumumab Following Intravenous Infusion in Relapsed or Refractory Multiple Myeloma After Prior Proteasome Inhibitor and Immunomodulatory Drug Treatment... Pamela L. Clemens, Xiaoyu Yan, Henk M. Lokhorst, Sagar Lonial, Nedjad Losic, Imran Khan, Richard Jansson, Tahamtan Ahmadi, Kristen Lantz, Honghui Zhou, 1 Thomas Puchalski, and Xu Steven Xu... Clin Pharmacokinet... 2017... Springer

Daratumumab depletes CD38+ immune regulatory cells, promotes T-cell expansion, and skews T-cell repertoire in multiple myeloma Jakub Krejcik, Tineke Casneuf, Inger S. Nijhof, Bie Verbist, Jaime Bald, Torben Plesner, Khaja Syed, Kevin Liu, Niels W. C. J. van de Donk, Brendan M. Weiss, Tahamtan Ahmadi, Henk M. Lokhorst, Tuna Mutis and A. Kate Sasser... Blood... 2016... Americal Society of Hematology

Direct in Vitro Comparison of Daratumumab with Surrogate Analogs of CD38 Antibodies MOR03087, SAR650984 and Ab79 Jeroen Lammerts van Bueren, Danielle Jakobs, Niels Kaldenhoven, Marcel Roza, Sanne Hiddingh, Joyce Meesters, Marleen Voorhorst, Elke Gresnigt, Luus Wiegman, Antonio Ortiz Buijsse, Grietje Andringa, Marije B. Overdijk, Parul Doshi, Kate Sasser, Michel de Weers and Paul W.H.I. Parren... Blood... 2014... Americal Society of Hematology

Daratumumab for the treatment of multiple myeloma... Touzeau C, Moreau P... Expert Opin Biol Ther... 2017... Taylor & Francis

Safety and efficacy of daratumumab in Japanese patients with relapsed or refractory multiple myeloma: a multicenter, phase 1, dose-escalation study... Iida S, Suzuki K, Kusumoto S, Ri M, Tsukada N, Abe Y, Aoki M, Inagaki M... Int J Hematol... 2017... Springer

How I manage the toxicities of myeloma drugs... Michel Delforge, Heinz Ludwig... Blood... 2017... Americal Society of Hematology

Daratumumab (anti-CD38) induces loss of CD38 on red blood cells... Harold C. Sullivan, Christian Gerner-Smidt, Ajay K. Nooka, Connie M. Arthur, Louisa Thompson, Amanda Mener, Seema R. Patel, Marianne Yee, Ross M. Fasano, Cassandra D. Josephson, Richard M. Kaufman, John D. Roback, Sagar Lonial and Sean R. Stowell... Blood... 2017... Americal Society of Hematology

**SCHEMATIC ASSAY PROCEDURE**


1. Remove all components, 30 minutes before adding into the assay plate.



2. Avoid repeated cool-thaw of the components as there will be a loss of activity and this can affect the results.




3.  Pipette **100 µl Standards / Samples** into the respective wells.

4.  Cover plate and incubate for  at 37°C.

5.  Aspirate and wash wells 4 times with **Wash Buffer (1X)**.

6.  Pipette **100 µl Daratumumab:HRP** into each well.

7.  Cover plate and incubate for  at 37°C.

8.  Aspirate and wash wells 4 times with **Wash Buffer (1X)**.

9.  Pipette **100 µl TMB Substrate** into each well.

10.  Cover plate and incubate for  at 37°C.

11.  Pipette **100 µl Stop Solution** into each well.

12. Read absorbance at 450nm with a  microplate reader within  of stopping reaction.

**Typical Example of a Work List**

Well #	Contents	Absorbance at 450nm	Mean Absorbance	ng/ml Anti-Daratumumab equivalent
1A 2A	zero std zero std			
1B 2B	10 ng/ml 10 ng/ml			
1C 2C	20 ng/ml 20 ng/ml			
1D 2D	40 ng/ml 40 ng/ml			
1E 2E	80 ng/ml 80 ng/ml			
1F 2F	160 ng/ml 160 ng/ml			
1G 2G	320 ng/ml 320 ng/ml			
1H 2H	640 ng/ml 640 ng/ml			
3A 4A	Sample			
3B 4B	Sample			

**LIMITED WARRANTY**

Krishgen Biosystems does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by Krishgen Biosystems, or against damages resulting from such non-Krishgen Biosystems made products or components. Krishgen Biosystems passes on to customer the warranty it received (if any) from the maker thereof of such non Krishgen made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Krishgen Biosystems.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of Krishgen Biosystems shall be to repair or replace the defective Products in the manner and for the period provided above. Krishgen Biosystems shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, and strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall Krishgen Biosystems be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of Krishgen Biosystems with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

Krishgen Biosystems. 2020

**THANK YOU FOR USING KRISHGEN PRODUCT!**

**SYMBOLS KEY**

<b>MTP</b>	Daratumumab Microtiter Plate (12 x 8 wells)
<b>STD</b>	Anti-Daratumumab Standard
<b>HRP CONJ</b>	Conjugate Horseradish Peroxidase
<b>SAMP DIL</b>	Sample Diluent
<b>20X WASH BUF</b>	(20X) Wash Buffer
<b>SUB TMB</b>	TMB Substrate
<b>SOLN STOP</b>	Stop Solution
	Consult Instructions for Use
<b>REF</b>	Catalog Number
	Expiration Date
	Storage Temperature

distributed in the US/Canada by:  
**EAGLE BIOSCIENCES, INC.**  
 20A NW Blvd, Suite 112 Nashua, NH 03063  
 Phone: 617-419-2019 FAX: 617-419-1110  
[www.EagleBio.com](http://www.EagleBio.com) • [info@eaglebio.com](mailto:info@eaglebio.com)



**EAGLE**  
BIOSCIENCES